

## WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

<b>CUSTOMER:</b>	Ash Waate
<b>SITE ADDRESS:</b>	Ellesmere Port
	Cheshire
<b>POST CODE:</b>	CH65 4LT
<b>WEIGHBRIDGE LOCATION:</b>	Out Weighbridge
<b>WEIGHBRIDGE SIZE:</b>	15x3

<b>REPORT NO.:</b>	4902650
<b>INDICATOR TYPE:</b>	I30
<b>INDICATOR SERIAL NO.:</b>	64F942564
<b>MAXIMUM CAPACITY (kg):</b>	50,000
<b>DIVISION 'e' (kg):</b>	20
<b>Weather Conditions</b>	Sunny

### REPEATABILITY TEST

Direction →  
of  
Travel ←

<b>Indicator Reading:</b>	43,280 kg
Weighbridge Viewed from Indicator Position	
<b>Indicator Reading:</b>	43,280 kg

Error Allowed  
+/-2dvs or +/-40kg  
**Pass**

### END - MIDDLE - END - ZERO TEST

<b>APPLIED LOAD (kg)</b>	10,000
--------------------------	--------

		End	Middle	End	Zero
<b>Fork Lift Truck Weight (kg)</b>	<b>INDICATOR READING (kg)</b>	9,980	10,000	9,980	0
<b>3,840</b>	<b>ERROR (e)</b>	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	<b>PASS/FAIL</b>	Pass	Pass	Pass	Pass

### RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-1dvs or 20kg
20,000	20,000	19,980	-1.0	Pass		error +/-2dvs or 40kg
Substitute Load	19,460	19,440	-1.0	Pass		error +/-2dvs or 40kg
25,000	24,460	24,440	-1.0	Pass	5,000	error +/-2dvs or 40kg
35,000	34,460	34,440	-1.0	Pass	10,000	error +/-2dvs or 40kg
40,000	39,460	39,440	-1.0	Pass	5,000	error +/-2dvs or 40kg
Substitute Load	38,300	38,280	-1.0	Pass	0	error +/-2dvs or 40kg
44,000	43,300	43,280	-1.0	Pass	5000	error +/-3dvs or 60kg

MAINT LOG UPDATED:

MAXIMUM RANGE TESTED:

TEST EQUIPMENT USED: Test Weights Traceable to National Standards

CALIBRATION DATE:

AUTHORISED ENGINEER: Steven Fox

TEST WEIGHT CERTIFICATE NUMBER: Tr12142Tr12143

NEXT CALIBRATION DATE:

### Comments

*We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction*

Name		Date	10/01/2024
Customer Signature		Engineer's Signature	

*We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.*